



PothHille

Registered in England No.82948
Established 1870

CERTIFICATE OF ANALYSIS

Product: Histowax with added Polymer

Batch: 519967

Date of Manufacture: 10.2019

Expiry Date: 10.2024

Poth Hille Specification No. PS9009

<u>Test</u>	<u>Method</u>	<u>Result</u>	<u>Specification</u>
Congealing Point (Rotating Thermometer)	ASTM D1321	55.0°C	130/135°F (54.0 – 57.0°C)
Visual	ASTM D156	+29	+25 min.
Flocculation		Pass	Pass

The above batch meets the product specification completely

Signed

..........

Matthew Cripps
QA Manager

Dated

.....16/10/2019.....